

1910.95 Appendix D - Audiometric test rooms

- * Standard Number: 1910.95 App D
- * Standard Title: Audiometric test rooms
- * SubPart Number: G
- * SubPart Title: Occupational Health and Environmental Control

Produced by USDOL OSHA - Directorate of Safety Standards &
Directorate of Health Standards

Maintained by USDOL OSHA - OCIS

This Appendix is Mandatory

Rooms used for audiometric testing shall not have background sound pressure levels exceeding those in Table D-1 when measured by equipment conforming at least to the Type 2 requirements of American National Standard Specification for Sound Level Meters, S1.4-1971 (R1976), and to the Class II requirements of American National Standard Specification for Octave, Half-Octave, and Third-Octave Band Filter Sets, S1.11-1971 (R1976).

TABLE D-1 - MAXIMUM ALLOWABLE OCTAVE-BAND SOUND PRESSURE LEVELS FOR AUDIOMETRIC TEST ROOMS

Octave-band center frequency (Hz).....	500	1000	2000	4000	8000
Sound pressure level (dB) ...	40	40	47	57	62
